

A Safe Workplace

A Workplace Safety and Health Manual for Your Community

Section: III-C	Rules Procedures, Practices and Guidelines	Page 1 of 1
		Approved By: Harvey Bostrom
Subject: G-19	Carbon Monoxide Airline Monitor Calibration	Effective: December 31, 2007
		Revised: April 1, 2011

General Requirements:

The carbon monoxide airline monitor should be calibrated by a trained technician prior to use or monthly if the unit is used on a regular basis.

Maintenance procedures requiring confined entry should be planned one month in advance to provide adequate time to have the unit calibrated.

In the event of an emergency situation, contact the Technical and Public Works Consultant assigned to your community for assistance with calibration and for a list of trained technicians.

Community Responsibilities

Council must ensure that:

- carbon monoxide airline monitor is calibrated prior to use; and
- equipment is used only by employees